

Channels	Part No.	Volume Range (µl)	Volume Increments (µl)	Test Volume (µl)	Accuracy (±%)	Precision (≤%)
24	6123	10 - 300	0.5	30	2.50	1.50
				150	1.50	1.00
				300	1.20	0.80
	6124	50 - 1250	1	125	2.50	1.50
				625	2.00	1.00
				1250	1.80	0.80
96	6101	0.5 - 12.5	0.01	1.25	8.00	3.00
				6.25	2.00	1.50
				12.5	1.50	1.00
	6106	2 - 50	0.05	5	5.00	2.50
				25	2.00	1.50
				50	1.50	1.00
	6102	5 - 125	0.1	12.5	3.00	1.70
				62.5	2.00	1.50
				125	1.50	1.00
	6103	10 - 300	0.5	30	2.50	1.50
				150	1.50	1.00
				300	1.20	0.80
	6104	50 - 1250	1	125	2.50	1.50
				625	2.00	1.00
				1250	1.80	0.80
384	6131	0.5 - 12.5	0.01	1.25	8.00	3.00
				6.25	2.00	1.50
				12.5	1.50	1.00
	6136	2 - 50	0.05	5	5.00	2.50
				25	2.00	1.50
				50	1.50	1.00
	6132	5 - 125	0.1	12.5	3.00	1.70
				62.5	2.00	1.50
				125	1.50	1.00

1. Specifications represent the performance of all channels.
2. Specifications apply to neat transfers in Pipet mode only.
3. Precision = coefficient of variation